

User instruction

The instruction below should be used in conjunction with detailed information on the packaging.

Short description of the product

Nitrile protective gloves, powder-free, non-sterile for disposable use. Product Reference "MERCATOR gogrip".

Full description of the product

| | |
|-----------------------|---|
| Raw material | : nitrile |
| External surface | : diamond textured |
| Internal surface | : polymerized + chlorinated |
| Cuff | : beaded |
| Colour | : orange |
| Shape | : ambidextrous, fitting to the right and left hand |
| Size range | : S (6-7), M (7-8), L (8-9), XL (9-10), XXL (10-11) |
| AQL | : 4.0 or customer required |
| Quantity in packaging | : 50 pcs. by weight |
| Shelf life | : 3 years (from the date of manufacturing) |


Storage instructions

It is recommended to store the gloves in dry place, in the temperature of 5-35°C and to protect them against direct sunlight and fluorescent light.

Keep the gloves in a distance of not less than 1m from heating devices, sources of fire and ozone.

Do not keep in direct vicinity of solvents, oils, fuels and lubricants.

Food contact

Gloves are marked with food contact symbol  and comply with the requirements of Commission Regulation (EU) No 10/2011, European Regulation (EC) No 1935/2004 and with Commission Regulation (EC) No 2023/2006 on Good Manufacturing Practice. Gloves are suitable for handling any type of food and have been tested for Overall Migration Test acc. DM 21/03/1973:

| Extraction conditions (tested for 2 h in 40°C) | Test Result (limit < 8 mg/dm ²) |
|--|---|
| 3% Acetic acid | Pass |
| 10% Ethanol | Pass |
| Olive oil | Pass |

PPE classification & compliance

Gloves are category III Personal Protective Equipment as per Annex II of the Regulation 2016/425 and comply to standards:

EN ISO 21420: 2020, EN ISO 374-1:2016+ A1:2018 (Type B), EN 374-2:2014, EN16523-1:2015, EN ISO 374-4:2019, EN ISO 374-5:2016.

EU Type Examination Certificate issued by: SATRA (Notified Body No. 2777)

UKCA Type Examination Certificate issued by: SATRA (Approved Body No. 0321)

Checking of PPE manufactured:

CE 2777
SATRA Technology Europe Limited.
Bractown Business Park.
Clonee.
D15YN2P.
Republic of Ireland.

UK
CA 0321
SATRA Technology Centre Limited.
Wyndham Way, Telford Way,
Kettering, Northamptonshire,
NN16 8SD,
United Kingdom.

Declaration of Conformity and Instruction of Use with the information about the importer are available at: www.mercatormedical.eu

Precautions and indications for use

Dry hands before putting the gloves on. Before usage, inspect the gloves for any defect or imperfections. These are disposable gloves. Do not let chemical substances get under the gloves through the cuff. If a chemical substance reaches the skin, wash it away immediately with plenty of water with soap. If the gloves get punctured, torn or broken during their use, take them off and put on the new ones. Avoid using gloves dirty in the inside as they may cause irritation leading to skin inflammation or more serious damages. The gloves should not be used in contact with open fire and to protect against any sharp tools. The gloves are not intended for welding, electric shock protection, ionizing radiation or from the effect of hot or cold objects.

This information does not reflect the actual duration of protection in the workplace and the between mixtures and pure chemicals.

Degradation results indicate the change in puncture resistance of the gloves after exposure to the challenge chemical. The chemical penetration resistance has been assessed under laboratory conditions from samples taken from the palm only (except in case where glove is equal to or over 400 mm – where the cuff is tested also) and relates only to the chemical tested and to the tested specimen. It can be different if the chemical is used in a mixture.

The penetration resistance has been assessed under laboratory conditions and relates only to the tested specimen. It is recommended to check that the gloves are suitable for the intended use because the conditions at the workplace may differ from the type test depending on the temperature, abrasion and degradation.

When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly. For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves.

Gloves are suitable for special purposes as they are protective gloves where risk of injury to the wrist is considered to be minimal, gloves are shorter than EN ISO 21420 min. length requirement.

Components / hazardous components

Some gloves may contain components known to be a possible cause of allergy for person allergic to them, who may develop contact irritation and/or allergic reaction. In case of an allergic reaction, seek medical assistance immediately.

Disposal

Used gloves can be contaminated with contagious or other hazardous substances. They should be disposed of in accordance with local regulation. Gloves should be buried or burned under controlled conditions.

Manufacturer

MERCATOR MEDICAL (Thailand) Ltd.
 88/8 Moo 12 Tambon Kampaengphet,
 Amphur Rattaphum, Songkhla 90180, Thailand.
www.mercator.co.th

Importer

Mercator Medical S.A.
 Address: ul. H. Modrzewskiej 30
 31-327 Krakow, Poland
www.mercatormedical.eu

Intended use

These are non-sterile protective gloves with thicker walls and diamond textured, providing excellent grip and protection when working in oily and wet environment.

These gloves are intended for use when there is a need for protection against dangerous substances and infective agents. These are chemical resistant gloves classified as Personal Protective Equipment category III. The design and labeling of these gloves correspond to the requirements of the Regulation 2016/425 on Personal Protective Equipment and included the PPE Regulation (EU) 2016/425 as brought into UK law and amended as a Category III product. Gloves should be used solely according to their intended purpose.
























| Permeation performance levels as per EN ISO 374-1:2016+ A1:2018 | | EN ISO 374-4:2019 Degradation [%] |
|---|--|-----------------------------------|
| Level1>10min | Level2>30 min, level3>60 min, level4>120min, Level5>240min, Level6>480m in | |
| Test results acc. to EN 16523-1:2015 | | |
| Chemical | Level | |
| n-Heptane (J) | 4 | 36.8 |
| 40% Sodium hydroxide (K) | 6 | -8.1 |
| 30% Hydrogen peroxide (P) | 3 | 31.9 |
| 37% Formaldehyde (T) | 4 | 14.8 |

Noted: 1) Glove minimum length for Lab application accordance to EN455-2


2) EN ISO 374-4: 2019 Degradation results indicate the change in puncture resistance of the gloves after exposure to the challenge chemical

| Test acc. To EN 374-2:2014 – Level 2 (ISO 2859) | | Test acc. To EN ISO 374-5:2016 | |
|---|--------|-------------------------------------|------|
| Performance level | AQL | Protection against bacteria & fungi | Pass |
| Level 3 | < 0.65 | Protection against viruses | Pass |
| Level 2 | <1.5 | | |
| Level 1 | < 4.0 | | |

Symbols used on the packaging

| | | | | | |
|---|--|---|---|--|--|
|  | Do not re-use / gloves are intended for single use |  | Non-sterile gloves |  | Powderfree gloves |
|  | Additional information on inner side of package |  | Keep away from solar and fluorescent light |  | Keep away from ozone |
|  | Do not use, if package is damaged |  | Temperature limitation / gloves store in temperature 5-35°C |  | 50 gloves by weight |
|  | Keep away from moisture, store in a dry place |  | Gloves protecting against chemical dangers with digit literal codes |  | Package made from paper, qualify for recycling |
|  | Protection against bacteria, fungi and viruses |  | Date of manufacture |  | Package is treated as municipal waste |
|  | Manufacturer, symbol should be accompanied by name and address of Manufacturer |  | Food contact symbol (article is suitable for food contact, for details check the instruction for use) |  | Lot / batch number |
|  | Expiry date |  | Indicates compliance with the requirements of Russian market |  | Indicates compliance with the requirements of Ukrainian market |
|  | Consult instructions for use |  | Catalogue number | | |

HOW TO PUT THE GLOVES ON?



HOW TO TAKE THE GLOVES OFF?

